



HydroSoft® 3D

血管内塞栓用コイル

Familiar DESIGN

- ・HyperSoft3Dと同じ
ループデザインで瘤内に均一に分布
- ・Primary windは0.012"～0.0135"
- ・内径0.0165"以上のマイクロカテーテルに適合
- ・SR付き

Familiar MATERIAL

- ・HydroSoftと同じハイドロゲルを使用した
第2世代ハイドロゲルコイル*1
- ・Reposition timeは30分、スチームプレップ不要
- ・蓄積された臨床エビデンス

Familiar DELIVERY

- ・V-Trak Advanced プッシャーの
コイルデリバリーシステムを採用
- ・シャフトと一体となったゴールド・コネクター
- ・シャフト長 195cm
- ・電気離脱式で、コイルのデタッチは0.75秒



症例を選ばずご使用いただける“Familiar coil”を目指して
MicroVentionのコイル技術を組み合わせました。

Easy



サイズダウンされた
最初のループが瘤内に留まります

Versatile



スペースの形状に合わせて内向き
にまとまりながら展開します

Deployment



最終ループまでの拡張力により、
コイルがスペース内で均一に分布します



図はイメージです

サイズ一覧

製品コード	2次ループ径 (mm)	コイル長 (cm)	JANコード	受注生産品
MV-00102HHTA	1	2	4987350006479	*
MV-00103HHTA	1	3	4987350006493	*
MV-01502HHTA	1.5	2	4987350006516	
MV-01503HHTA	1.5	3	4987350006530	
MV-01504HHTA	1.5	4	4987350006554	*
MV-00202HHTA	2	2	4987350006578	
MV-00203HHTA	2	3	4987350006295	
MV-00204HHTA	2	4	4987350006592	
MV-00206HHTA	2	6	4987350006615	
MV-00208HHTA	2	8	4987350006370	
MV-02504HHTA	2.5	4	4987350006639	
MV-02506HHTA	2.5	6	4987350006653	
MV-02508HHTA	2.5	8	4987350006677	*
MV-00304HHTA	3	4	4987892028977	
MV-00306HHTA	3	6	4987350006691	
MV-00308HHTA	3	8	4987350006714	
MV-00310HHTA	3	10	4987350006738	*
MV-00408HHTA	4	8	4987892070174	
MV-00410HHTA	4	10	4987892070198	
MV-00412HHTA	4	12	4987892070211	
MV-00510HHTA	5	10	4987892070259	
MV-00515HHTA	5	15	4987892070273	
MV-00612HHTA	6	12	4987892070310	
MV-00619HHTA	6	19	4987892070334	
MV-00715HHTA	7	15	4987892070372	
MV-00728HHTA	7	28	4987892070396	
MV-00817HHTA	8	17	4987892070433	
MV-00833HHTA	8	33	4987892070457	

*…受注生産品の在庫につきましては、弊社担当者までお問合せください。

■ 第2世代ハイドロゲルコイル関連文献

1. Taschner, C. GREAT – Second Generation Hydrogel-coated Coils for the Endovascular Treatment of Intracranial Aneurysms: Final Results of a Randomized Controlled Trial. 2017 International Stroke Conference Abstract Presentation.
2. Brinjikji, W. GEL-THE-NEC: a prospective registry evaluating the safety, ease of use, and efficacy of the HydroSoft coil as a finishing device. Journal of NeuroInterventional Surgery 2017; 0: 1-5 (Online first)
3. Dabus, G. Safety, immediate and mid-term results of the newer generation of hydrogel coils in the treatment of ruptured aneurysms; a multicenter study. Journal of NeuroInterventional Surgery 2016; 9(4): 419-424
4. Park, JH. Embolization of Intracranial Aneurysms with HydroSoft Coils: Results of Korean Multicenter Study. American Journal of Neuroradiology 2011; 32: 1756-1761

一般的名称: 中心循環系血管内塞栓促進用補綴材 販売名: V-Trak/ハイドロソフトエンボリックシステム 医療機器承認番号: 22400BZX00016

※本製品の詳細は、添付文書をご参照ください。

